

## TECHNICAL DATA

# Fluke 787 ProcessMeter Digital Multimeter



## Key features

- DMM and a loop calibrator in one rugged, handheld tool
- Provides a precision 1000 V, 440 mA True-RMS digital multimeter designed to meet 1000 volt IEC 1010 CAT III standards
- Features a clear LCD with backlight; 4,000 counts (30,000 counts for DC current)
- Performs frequency measurements to 20 kHz

## Product overview: Fluke 787 ProcessMeter Digital Multimeter

### Fluke 787 ProcessMeter doubles your capabilities but not your cost

The Fluke 787 ProcessMeter combines a DMM and a Loop Calibrator in one rugged, handheld tool, for about what you would expect to pay for a loop calibrator alone. It's a maintenance and calibration tool that will be at home in every instrumentation technician's tool box. Based on the trusted Fluke 87 DMM, the 787 adds the ability to measure, source, and simulate dc loop current with 1 microamp resolution and 0.05% accuracy. It puts a complete solution for troubleshooting and calibrating current loop applications in the palm of your hand so you can get more information, faster and easier.

So if you're tired of having to track down a calibrator, lugging around two tools, and switching back and forth, get the convenient solution for doubling your power. The 787 ProcessMeter from Fluke.

Other useful features:

- Provides Min/Max/Average/Hold/Relative modes, and diode test and continuity beeper
- Displays simultaneous mA and % of scale readout
- Provides 20 mA DC current source/loop calibrator/simulator
- Manual step (100%, 25%, Coarse, Fine) plus auto step and auto ramp
- Externally accessible battery for easy battery changes
- 1000 V overload protection on V, ohms, frequency
- 150 V overload protection on mA, backed up by 440 mA 1000V fuse

**Want a built in 24-Volt loop supply and 250 Ω HART resistor? Consider the Fluke 789.**

## Specifications: Fluke 787 ProcessMeter Digital Multimeter

Specifications		
Current measurement	<b>Range</b>	0-1 A
	Resolution	1 mA
	Accuracy	0.2% + 2 LSD
	<b>Range</b>	0-30 mA
	Resolution	0.001 mA
	Accuracy	0.05% + 2 LSD
Current sourcing	<b>Range</b>	0-20 mA or 4-20 mA
	Resolution	0.05% of span
	Drive Capability	500 u03a9 @ 24 mA

Frequency	To 19.999 kHz, 0.005% + 1 LSD	
Continuity	Beeps for resistance < 100 ohms	
Voltage measurement	<b>Range</b>	0-1000 V AC or DC
	Resolution	0.1 mV to 1.0 V
	Accuracy	0.1% Rdg + 1 LSD (VDC)
Resistance measurement	To 40 Mu03a9, 0.2% + 1 LSD	
Diode test	2.4 V shows diode voltage drop	
<b>Environmental Specifications</b>		
Operating temperature	-20u00b0C to 55u00b0C	
Storage temperature	-40u00b0C to 60u00b0C	
Humidity (without condensation)	95% up to 30u00b0C	
	75% up to 40u00b0C	
	45% up to 50u00b0C	
	35% up to 55u00b0C	
Temperature coefficient	0.05 x (specified accuracy) per u00b0C (for temperatures < 18u00b0C or > 28u00b0C)	
<b>Safety Specifications</b>		
Electrical safety	Designed in accordance with IEC 1010-1, ANSI/ISA S82.01-1994 and CAN/CSA C22.2 No. 1010.1-92	
Overvoltage category	IEC 1010-1 CAT III	
Maximum voltage	1000 V rms	
<b>Mechanical and General Specifications</b>		
Size	32 x 87 x 187 mm	
Weight	369 g	
Baeries	Single 9 V Alkaline baery (ANSI/NEDA 1604A or IEC 6LR61)	
Warranty	3 years	
Baery life	50 hours typical (measurement), 12 hours typical (sourcing 12 mA)	
Shock and vibration	Random, 2 g, 5-500 Hz	
	1 meter drop test	
Display current and % of span	Yes	
Auto step, auto ramp	Yes	

## Ordering information



### **Fluke 787**

Fluke 787 ProcessMeter™

Includes:

- AC70A Alligator Clips
- C81Y Protective Yellow Holster with Flex-Stand™
- One 9V Alkaline Battery
- Quick Reference Guide
- TL75 Safety-Designed Test lead Set



**Fluke.** *Keeping your world up and running.®*

**Fluke Corporation**  
PO Box 9090, Everett, WA 98206 U.S.A.

**For more information call:**  
In the U.S.A. (800) 443-5853  
In Canada (800) 36-FLUKE  
From other countries +1 (425) 446-5500  
[www.fluke.com](http://www.fluke.com)

©2025 Fluke Corporation.  
Specifications subject to change without notice.  
01/2025

**Modification of this document is not permitted  
without written permission from Fluke Corporation.**